

ORIGINAL

**Calibrating Unit
New Standard Solution Report**

*Sol #169
3/10/15*

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01521
Certification File No.: 01522
Linearity File No.: 01523
Solution File No.: 01595
Sequential File No.: 01595

Serial No.: ARAJ-0075

Calib. Date: 11/05/2014
Cert. Date: 11/05/2014
Lin. Date: 11/05/2014
Soln. Date: 03/10/2015
File Date: 03/10/2015

Calib. No.: 00017
Cert. No.: 00012
Lin. No.: 00012
Soln. No.: 00169

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 14F130

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 06/20/2016
Bottle No.: 0022

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:05S		
Control 1 EC	0.100%	15:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06S		
Control 2 EC	0.098%	15:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:08S		
Control 3 EC	0.099%	15:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: Sgt. J. Em #282

Badge No.: 1282

Date: 03/10/2015