Calibrating Unit
New Standard Solution Report

SOL # 170
4-7-15

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.				
Calibration File No.:	01604	Calib. Date:	04/07/2015	Calib. No.:	00018
Certification File No.:	01605	Cert. Date:	04/07/2015	Cert. No.:	00013
Linearity File No.:	01606	Lin. Date:	04/07/2015	Lin. No.:	00013
Solution File No.:	01607	Soln. Date:	04/07/2015	Soln. No.:	00170
Sequential File No.:	01607	File Date:	04/07/2015		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAN-0041
Control Solution %:	0.100%			Expires:	06/20/2016
Solution Control Lot:	14F130			Bottle No.:	0016

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:38D		
Control 1 EC	0.101%	13:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39D		
Control 2 EC	0.100%	13:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:40D		
Control 3 EC	0.100%	13:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-436

Changed By:

Last Name: DELLANOCE First Name: JOSEPH MI: S

Signature: SGT. [signature] # 6027

Badge No.: 6027

Date: 04/07/2015