

ORIGINAL

Calibrating Unit New Standard Solution Report

Solo # 171

4/22/15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01604
Certification File No.: 01605
Linearity File No.: 01606
Solution File No.: 01616
Sequential File No.: 01616

Serial No.: ARAJ-0075
Calib. Date: 04/07/2015
Cert. Date: 04/07/2015
Lin. Date: 04/07/2015
Soln. Date: 04/22/2015
File Date: 04/22/2015

Calib. No.: 00018
Cert. No.: 00013
Lin. No.: 00013
Soln. No.: 00171

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 14F130

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 06/20/2016
Bottle No.: 0021

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 15:04D | | |
| Control 1 EC | 0.100% | 15:05D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 15:05D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:06D | | |
| Control 2 EC | 0.098% | 15:06D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 15:06D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:07D | | |
| Control 3 EC | 0.099% | 15:08D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 15:08D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 15:08D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: *P.H. Kenneth Thomas #314*

Badge No.: 314

Date: 04/22/2015