

Calibrating Unit
New Standard Solution Report

ORIGINAL

Sol. # 200

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01254
Certification File No.: 01255
Linearity File No.: 01256
Solution File No.: 01261
Sequential File No.: 01261

Calib. Date: 04/07/2015
Cert. Date: 04/07/2015
Lin. Date: 04/07/2015
Soln. Date: 05/13/2015
File Date: 05/13/2015

Serial No.: ARWF-0394

Calib. No.: 00029
Cert. No.: 00021
Lin. No.: 00021
Soln. No.: 00200

5/13/15
Failed d/oll

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 14B127

Model No.: CU-34

Serial No.: DDAN-0040
Expires: 02/14/2016
Bottle No.: 0816

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:39D		
Control 1 EC	0.094%	09:40D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.099%	09:40D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	09:41D		
Control 2 EC	0.093%	09:41D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.099%	09:41D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	09:42D		
Control 3 EC	0.092%	09:43D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.097%	09:43D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	09:44D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____

Badge No.: 1282

Date: 05/13/2015