

Calibrating Unit

New Standard Solution Report

ORIGINAL

SOLO #202

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01254
 Certification File No.: 01255
 Linearity File No.: 01256
 Solution File No.: 01263
 Sequential File No.: 01263

Serial No.: ARWF-0394

Calib. Date: 04/07/2015
 Cert. Date: 04/07/2015
 Lin. Date: 04/07/2015
 Soln. Date: 05/13/2015
 File Date: 05/13/2015

Calib. No.: 00029
 Cert. No.: 00021
 Lin. No.: 00021
 Soln. No.: 00202

5/13/15
 Failed (M)

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0040
 Expires: 02/11/2017
 Bottle No.: 0921

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:28D		
Control 1 EC	0.095%	14:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:29D		
Control 2 EC	0.093%	14:30D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.097%	14:30D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:31D		
Control 3 EC	0.093%	14:32D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.097%	14:32D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:33D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____

J. E. #282

Badge No.: 1282

Date: 05/13/2015