

ORIGINAL

# Calibrating Unit New Standard Solution Report

Sol. # 176  
6/26/15

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01604  
Certification File No.: 01605  
Linearity File No.: 01606  
Solution File No.: 01663  
Sequential File No.: 01663

Calib. Date: 04/07/2015  
Cert. Date: 04/07/2015  
Lin. Date: 04/07/2015  
Soln. Date: 06/26/2015  
File Date: 06/26/2015

Serial No.: ARAJ-0075  
Calib. No.: 00018  
Cert. No.: 00013  
Lin. No.: 00013  
Soln. No.: 00176

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 02/11/2017  
Bottle No.: 0923

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:36D		
Control 1 EC	0.098%	22:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:37D		
Control 2 EC	0.098%	22:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:39D		
Control 3 EC	0.098%	22:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: \_\_\_\_\_

*Kenneth J Thomas #314*

Badge No.: 314

Date: 06/26/2015