

ORIGINAL

Calibrating Unit New Standard Solution Report

Sol. # 177
7/4/15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01604
Certification File No.: 01605
Linearity File No.: 01606
Solution File No.: 01669
Sequential File No.: 01669

Serial No.: ARAJ-0075
Calib. Date: 04/07/2015
Cert. Date: 04/07/2015
Lin. Date: 04/07/2015
Soln. Date: 07/04/2015
File Date: 07/04/2015
Calib. No.: 00018
Cert. No.: 00013
Lin. No.: 00013
Soln. No.: 00177

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0920

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 16:47D | | |
| Control 1 EC | 0.099% | 16:47D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.096% | 16:47D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:48D | | |
| Control 2 EC | 0.097% | 16:49D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 16:49D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:49D | | |
| Control 3 EC | 0.097% | 16:50D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 16:50D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:50D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth Thomas
#314

Badge No.: 314

Date: 07/04/2015