

ORIGINAL

Calibrating Unit New Standard Solution Report

SOL. # 178

7/9/15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01604
Certification File No.: 01605
Linearity File No.: 01606
Solution File No.: 01674
Sequential File No.: 01674

Serial No.: ARAJ-0075
Calib. Date: 04/07/2015 **Calib. No.:** 00018
Cert. Date: 04/07/2015 **Cert. No.:** 00013
Lin. Date: 04/07/2015 **Lin. No.:** 00013
Soln. Date: 07/09/2015 **Soln. No.:** 00178
File Date: 07/09/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34 **Serial No.:** DDAN-0041
Expires: 02/11/2017
Bottle No.: 0918

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 16:59D | | |
| Control 1 EC | 0.098% | 17:00D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 17:00D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:01D | | |
| Control 2 EC | 0.097% | 17:01D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 17:01D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:02D | | |
| Control 3 EC | 0.097% | 17:02D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 17:02D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:03D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature:  #314

Badge No.: 314

Date: 07/09/2015