

Calibrating Unit

New Standard Solution Report

ORIGINAL

Sol. 182

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01604
Certification File No.: 01605
Linearity File No.: 01606
Solution File No.: 01700
Sequential File No.: 01700

Serial No.: ARAJ-0075

8/14/15

Calib. Date: 04/07/2015
Cert. Date: 04/07/2015
Lin. Date: 04/07/2015
Soln. Date: 08/14/2015
File Date: 08/14/2015

Calib. No.: 00018
Cert. No.: 00013
Lin. No.: 00013
Soln. No.: 00182

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0916

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:29D		
Control 1 EC	0.098%	23:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:30D		
Control 2 EC	0.098%	23:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:32D		
Control 3 EC	0.098%	23:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: _____



Badge No.: 289

Date: 08/14/2015