

**ORIGINAL**

**Calibrating Unit  
New Standard Solution Report**

*Sol. # 207  
8/24/15*

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01266  
Certification File No.: 01267  
Linearity File No.: 01268  
Solution File No.: 01305  
Sequential File No.: 01305

Serial No.: ARWF-0394  
Calib. No.: 00031  
Cert. No.: 00022  
Lin. No.: 00022  
Soln. No.: 00207

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15B137

Model No.: CU-34  
Serial No.: DDAN-0040  
Expires: 02/11/2017  
Bottle No.: 0913

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:13D		
Control 1 EC	0.101%	15:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14D		
Control 2 EC	0.100%	15:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:16D		
Control 3 EC	0.100%	15:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAIRE

First Name: GUY

MI: E

Signature: \_\_\_\_\_

*JEM #282*

Badge No.: 1282

Date: 08/24/2015