

# Calibrating Unit

## New Standard Solution Report

ORIGINAL

Sol. # 208

9/28/15

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01266  
Certification File No.: 01267  
Linearity File No.: 01268  
Solution File No.: 01323  
Sequential File No.: 01323

Serial No.: ARWF-0394

Calib. Date: 06/10/2015  
Cert. Date: 06/10/2015  
Lin. Date: 06/10/2015  
Soln. Date: 09/28/2015  
File Date: 09/28/2015

Calib. No.: 00031  
Cert. No.: 00022  
Lin. No.: 00022  
Soln. No.: 00208

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0040  
Expires: 02/11/2017  
Bottle No.: 0905

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:37D		
Control 1 EC	0.102%	09:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 2 EC	0.101%	09:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 3 EC	0.100%	09:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: \_\_\_\_\_



Badge No.: 1282

Date: 09/28/2015