

Sol. 189
10-2-15

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01746
Certification File No.: 01747
Linearity File No.: 01748
Solution File No.: 01749
Sequential File No.: 01749

Calib. Date: 10/02/2015
Cert. Date: 10/02/2015
Lin. Date: 10/02/2015
Soln. Date: 10/02/2015
File Date: 10/02/2015

Serial No.: ARAJ-0075
Calib. No.: 00019
Cert. No.: 00014
Lin. No.: 00014
Soln. No.: 00189

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0903

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:09D		
Control 1 EC	0.101%	12:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.100%	12:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		
Control 3 EC	0.099%	12:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXK P2 - 436

Changed By:

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: SGT. [Signature] # 6027

Badge No.: 6027

Date: 10/02/2015