

Calibrating Unit New Standard Solution Report

Sol. # 209

10-2-15

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01324 Calib. Date: 10/02/2015 Calib. No.: 00032
Certification File No.: 01325 Cert. Date: 10/02/2015 Cert. No.: 00023
Linearity File No.: 01326 Lin. Date: 10/02/2015 Lin. No.: 00023
Solution File No.: 01327 Soln. Date: 10/02/2015 Soln. No.: 00209
Sequential File No.: 01327 File Date: 10/02/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 02/11/2017
Solution Control Lot: 15B137 Bottle No.: 0906

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:05D		
Control 1 EC	0.100%	13:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:07D		
Control 2 EC	0.100%	13:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08D		
Control 3 EC	0.100%	13:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-444 *jso*

Changed By:

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: SGT. JSP # 6027

Badge No.: 6027

Date: 10/02/2015