

Calibrating Unit New Standard Solution Report

ORIGINAL

Sol # 191

10-29-15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01746
Certification File No.: 01747
Linearity File No.: 01748
Solution File No.: 01763
Sequential File No.: 01763

Serial No.: ARAJ-0075
Calib. Date: 10/02/2015
Cert. Date: 10/02/2015
Lin. Date: 10/02/2015
Soln. Date: 10/29/2015
File Date: 10/29/2015

Calib. No.: 00019
Cert. No.: 00014
Lin. No.: 00014
Soln. No.: 00191

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 02/11/2017
Bottle No.: 0909

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 04:18S | | |
| Control 1 EC | 0.099% | 04:18S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 04:18S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:19S | | |
| Control 2 EC | 0.098% | 04:20S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 04:20S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:20S | | |
| Control 3 EC | 0.097% | 04:21S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 04:21S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:22S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: _____



Badge No.: 289

Date: 10/29/2015