

ORIGINAL

Calibrating Unit New Standard Solution Report

501.#
210
11-6-15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01324
Certification File No.: 01325
Linearity File No.: 01326
Solution File No.: 01330
Sequential File No.: 01330

Serial No.: ARWF-0394
Calib. No.: 00032
Cert. No.: 00023
Lin. No.: 00023
Soln. No.: 00210
File Date: 11/06/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0040
Expires: 02/11/2017
Bottle No.: 0907

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:11S		
Control 1 EC	0.100%	13:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13S		
Control 2 EC	0.100%	13:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14S		
Control 3 EC	0.099%	13:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: Kenneth Thomas #314

Badge No.: 314

Date: 11/06/2015