

# Calibrating Unit New Standard Solution Report

**ORIGINAL**

Sol. # 192

11/13/15

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01746  
 Certification File No.: 01747  
 Linearity File No.: 01748  
 Solution File No.: 01769  
 Sequential File No.: 01769

Serial No.: ARAJ-0075  
 Calib. Date: 10/02/2015  
 Cert. Date: 10/02/2015  
 Lin. Date: 10/02/2015  
 Soln. Date: 11/13/2015  
 File Date: 11/13/2015

Calib. No.: 00019  
 Cert. No.: 00014  
 Lin. No.: 00014  
 Soln. No.: 00192

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 15B137

Model No.: CU-34

Serial No.: DDAN-0041  
 Expires: 02/11/2017  
 Bottle No.: 0901

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:38S		
Control 1 EC	0.099%	04:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:39S		
Control 2 EC	0.097%	04:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:41S		
Control 3 EC	0.097%	04:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SYSOL

First Name: TIM

MI: M

Signature: \_\_\_\_\_

*Handwritten Signature* 289

Badge No.: 289

Date: 11/13/2015