

Calibrating Unit

New Standard Solution Report

ORIGINAL

Sol. 211

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01324
 Certification File No.: 01325
 Linearity File No.: 01326
 Solution File No.: 01339
 Sequential File No.: 01339

Serial No.: ARWF-0394

Calib. Date: 10/02/2015
 Cert. Date: 10/02/2015
 Lin. Date: 10/02/2015
 Soln. Date: 12/10/2015
 File Date: 12/10/2015

Calib. No.: 00032
 Cert. No.: 00023
 Lin. No.: 00023
 Soln. No.: 00211

12/10/15

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0040
 Expires: 04/13/2017
 Bottle No.: 0202

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:38S		
Control 1 EC	0.100%	01:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:40S		
Control 2 EC	0.100%	01:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:41S		
Control 3 EC	0.099%	01:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: _____



Badge No.: 289

Date: 12/10/2015