

# Calibrating Unit New Standard Solution Report

ORIGINAL  
Sol. #196

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01746  
Certification File No.: 01747  
Linearity File No.: 01748  
Solution File No.: 01788  
Sequential File No.: 01788

Serial No.: ARAJ-0075

12/13/15

Calib. Date: 10/02/2015  
Cert. Date: 10/02/2015  
Lin. Date: 10/02/2015  
Soln. Date: 12/13/2015  
File Date: 12/13/2015

Calib. No.: 00019  
Cert. No.: 00014  
Lin. No.: 00014  
Soln. No.: 00196

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 04/13/2017  
Bottle No.: 0207

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:57S		
Control 1 EC	-.---%	03:58S	22.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	03:58S	22.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: \_\_\_\_\_

 289

Badge No.: 289

Date: 12/13/2015