

Calibrating Unit
New Standard Solution Report

[ORIGINAL

SOI. H. 197

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01746
Certification File No.: 01747
Linearity File No.: 01748
Solution File No.: 01789
Sequential File No.: 01789

Serial No.: ARAJ-0075 12/13/15

Calib. Date: 10/02/2015 Calib. No.: 00019
Cert. Date: 10/02/2015 Cert. No.: 00014
Lin. Date: 10/02/2015 Lin. No.: 00014
Soln. Date: 12/13/2015 Soln. No.: 00197
File Date: 12/13/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34 Serial No.: DDAN-0041
Expires: 04/13/2017
Bottle No.: 0207

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:58S		
Control 1 EC	0.097%	08:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 2 EC	0.097%	09:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		
Control 3 EC	0.096%	09:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SYSOL

First Name: TIM

MI: M

Signature: _____

T Sysol #289

Badge No.: 289

Date: 12/13/2015