

Journal

# Calibrating Unit New Standard Solution Report

Sol. # 199  
12-24-15

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01746  
Certification File No.: 01747  
Linearity File No.: 01748  
Solution File No.: 01799  
Sequential File No.: 01799

Serial No.: ARAJ-0075

Calib. Date: 10/02/2015  
Cert. Date: 10/02/2015  
Lin. Date: 10/02/2015  
Soln. Date: 12/24/2015  
File Date: 12/24/2015

Calib. No.: 00019  
Cert. No.: 00014  
Lin. No.: 00014  
Soln. No.: 00199

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 04/13/2017  
Bottle No.: 0213

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:21S		
Control 1 EC	0.099%	19:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:23S		
Control 2 EC	0.098%	19:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:24S		
Control 3 EC	0.098%	19:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: \_\_\_\_\_

*Kenneth Thomas* #314

Badge No.: 314

Date: 12/24/2015