

Calibrating Unit New Standard Solution Report

ORIGINAL

Sol. #

213

1/9/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01324
Certification File No.: 01325
Linearity File No.: 01326
Solution File No.: 01357
Sequential File No.: 01357

Calib. Date: 10/02/2015
Cert. Date: 10/02/2015
Lin. Date: 10/02/2015
Soln. Date: 01/09/2016
File Date: 01/09/2016

Serial No.: ARWF-0394
Calib. No.: 00032
Cert. No.: 00023
Lin. No.: 00023
Soln. No.: 00213

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0040
Expires: 04/13/2017
Bottle No.: 0204

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:45S		
Control 1 EC	0.100%	17:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:47S		
Control 2 EC	0.100%	17:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:48S		
Control 3 EC	0.100%	17:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAIRE

First Name: SGT. GUY

MI: -

Signature: _____

J. GUY #282

Badge No.: 282

Date: 01/09/2016