

# Calibrating Unit

## New Standard Solution Report

ORIGINAL

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01746  
 Certification File No.: 01747  
 Linearity File No.: 01748  
 Solution File No.: 01813  
 Sequential File No.: 01813

Serial No.: ARAJ-0075  
 Calib. Date: 10/02/2015  
 Cert. Date: 10/02/2015  
 Lin. Date: 10/02/2015  
 Soln. Date: 01/31/2016  
 File Date: 01/31/2016

Sol #  
203  
2-2-16

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDAN-0041  
 Expires: 04/13/2017  
 Bottle No.: 0220

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:21S		
Control 1 EC	0.099%	07:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23S		
Control 2 EC	0.097%	07:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24S		
Control 3 EC	0.097%	07:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SYSOL

First Name: TIM

MI: M

Signature: \_\_\_\_\_

*Handwritten signature: Sysol 289*

Badge No.: 289

Date: 01/31/2016