

# Calibrating Unit

## New Standard Solution Report

80/215

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01324 Calib. Date: 10/02/2015 Calib. No.: 00032  
Certification File No.: 01325 Cert. Date: 10/02/2015 Cert. No.: 00023  
Linearity File No.: 01326 Lin. Date: 10/02/2015 Lin. No.: 00023  
Solution File No.: 01374 Soln. Date: 02/09/2016 Soln. No.: 00215  
Sequential File No.: 01374 File Date: 02/09/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387  
Control Solution %: 0.100% Expires: 04/13/2017  
Solution Control Lot: 15D138 Bottle No.: 0218

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:40S		
Control 1 EC	0.100%	13:41S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42S		
Control 2 EC	0.100%	13:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:43S		
Control 3 EC	0.102%	13:44S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔWJpz - 026 (KA)

### Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpr. I K W. G #6704

Badge No.: 6704

Date: 02/09/2016