

Calibrating Unit

New Standard Solution Report

Sol. # 205
3-18-16

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01818
 Certification File No.: 01819
 Linearity File No.: 01820
 Solution File No.: 01821
 Sequential File No.: 01821

Serial No.: ARAJ-0075

Calib. Date: 03/18/2016 Calib. No.: 00021
 Cert. Date: 03/18/2016 Cert. No.: 00015
 Lin. Date: 03/18/2016 Lin. No.: 00015
 Soln. Date: 03/18/2016 Soln. No.: 00205
 File Date: 03/18/2016

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDWK S3-0386
 Expires: 04/13/2017
 Bottle No.: 0221

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:11S	33.9°C	*** TEST PASSED ***
Control 1 EC	0.101%	10:11S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12S	33.9°C	*** TEST PASSED ***
Control 2 EC	0.101%	10:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13S	33.9°C	*** TEST PASSED ***
Control 3 EC	0.101%	10:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14S	33.9°C	*** TEST PASSED ***

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-025 (KA)

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TP. I K W. 9 #6704

Badge No.: 6704

Date: 03/18/2016