

# Calibrating Unit New Standard Solution Report

Sol. # 220  
3/18/16

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01407  
 Certification File No.: 01408  
 Linearity File No.: 01409  
 Solution File No.: 01410  
 Sequential File No.: 01410

Calib. Date: 03/18/2016  
 Cert. Date: 03/18/2016  
 Lin. Date: 03/18/2016  
 Soln. Date: 03/18/2016  
 File Date: 03/18/2016

Serial No.: ARWF-0394  
 Calib. No.: 00033  
 Cert. No.: 00024  
 Lin. No.: 00024  
 Soln. No.: 00220

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 15D138

Model No.: CU-34

Serial No.: DDWK S3-0387  
 Expires: 04/13/2017  
 Bottle No.: 0211

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:43S	13:43 Δ KA	
Control 1 EC	0.100%	12:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:45S		
Control 2 EC	0.100%	12:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:46S		
Control 3 EC	0.100%	12:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47S	13:47 Δ KA	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔWJTP2-026 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpr. I RK W. G #6704

Badge No.: 6704

Date: 03/18/2016