

Calibrating Unit New Standard Solution Report

Sol. # 221
4/2/16

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01407	Calib. Date: 03/18/2016
Certification File No.:	01408	Cert. No.: 00033
Linearity File No.:	01409	Cert. Date: 03/18/2016
Solution File No.:	01420	Lin. No.: 00024
Sequential File No.:	01420	Soln. Date: 04/02/2016
		Soln. No.: 00221
		File Date: 04/02/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 0224

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:46S		
Control 1 EC	0.099%	21:47S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:47S		
Control 2 EC	0.099%	21:48S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:49S		
Control 3 EC	0.099%	21:49S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:50S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: 

Badge No.: 314

Date: 04/02/2016