



**ORIGINAL**

# Calibrating Unit New Standard Solution Report

Sol. #224

6/4/16

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01407  
 Certification File No.: 01408  
 Linearity File No.: 01409  
 Solution File No.: 01452  
 Sequential File No.: 01452

Calib. Date: 03/18/2016  
 Cert. Date: 03/18/2016  
 Lin. Date: 03/18/2016  
 Soln. Date: 06/04/2016  
 File Date: 06/04/2016

Serial No.: ARWF-0394  
 Calib. No.: 00033  
 Cert. No.: 00024  
 Lin. No.: 00024  
 Soln. No.: 00224

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDWK S3-0387  
 Expires: 03/22/2018  
 Bottle No.: 0804

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 18:19D        |                               |                           |
| Control 1 EC      | 0.100%         | 18:20D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.101%         | 18:20D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 18:20D        |                               |                           |
| Control 2 EC      | 0.099%         | 18:21D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.101%         | 18:21D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 18:21D        |                               |                           |
| Control 3 EC      | 0.101%         | 18:22D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 18:22D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 18:23D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: \_\_\_\_\_

*Kenneth Thom* #314

Badge No.: 314

Date: 06/04/2016