



ORIGINAL

Calibrating Unit New Standard Solution Report

Sol. #
209
6/11/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01818
Certification File No.: 01819
Linearity File No.: 01820
Solution File No.: 01868
Sequential File No.: 01868

Calib. Date: 03/18/2016
Cert. Date: 03/18/2016
Lin. Date: 03/18/2016
Soln. Date: 06/11/2016
File Date: 06/11/2016

Serial No.: ARAJ-0075
Calib. No.: 00021
Cert. No.: 00015
Lin. No.: 00015
Soln. No.: 00209

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 03/22/2018
Bottle No.: 0413

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:47D		
Control 1 EC	0.100%	23:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:48D		
Control 2 EC	0.098%	23:48D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:49D		
Control 3 EC	0.099%	23:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth Thom #314

Badge No.: 314

Date: 06/11/2016