

Calibrating Unit New Standard Solution Report

SOI # 212
ORIGINAL

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01818
Certification File No.: 01819
Linearity File No.: 01820
Solution File No.: 01886
Sequential File No.: 01886

Serial No.: ARAJ-0075

Calib. Date: 03/18/2016
Cert. Date: 03/18/2016
Lin. Date: 03/18/2016
Soln. Date: 06/29/2016
File Date: 06/29/2016

Calib. No.: 00021
Cert. No.: 00015
Lin. No.: 00015
Soln. No.: 00212

6/29/16

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 03/22/2018
Bottle No.: 0807

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:27D		
Control 1 EC	0.099%	19:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:28D		
Control 2 EC	0.098%	19:29D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:29D		
Control 3 EC	0.098%	19:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 06/29/2016