

# Calibrating Unit New Standard Solution Report



**ORIGINAL**

*Sol. # 213*

*7/12/16*

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01818  
 Certification File No.: 01819  
 Linearity File No.: 01820  
 Solution File No.: 01894  
 Sequential File No.: 01894

Serial No.: ARAJ-0075  
 Calib. Date: 03/18/2016  
 Cert. Date: 03/18/2016  
 Lin. Date: 03/18/2016  
 Soln. Date: 07/12/2016  
 File Date: 07/12/2016

Calib. No.: 00021  
 Cert. No.: 00015  
 Lin. No.: 00015  
 Soln. No.: 00213

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDWK S3-0386  
 Expires: 03/22/2018  
 Bottle No.: 0805

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 12:01D |                |                     |
| Control 1 EC      | 0.099% | 12:02D | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 12:02D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:02D |                |                     |
| Control 2 EC      | 0.099% | 12:03D | 33.9°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 12:03D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:03D |                |                     |
| Control 3 EC      | 0.099% | 12:04D | 33.9°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 12:04D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:05D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: 

Badge No.: 306

Date: 07/12/2016