



ORIGINAL

Calibrating Unit

New Standard Solution Report

SOL-214
7-23-16

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01818
 Certification File No.: 01819
 Linearity File No.: 01820
 Solution File No.: 01902
 Sequential File No.: 01902

Serial No.: ARAJ-0075

Calib. Date: 03/18/2016 Calib. No.: 00021
 Cert. Date: 03/18/2016 Cert. No.: 00015
 Lin. Date: 03/18/2016 Lin. No.: 00015
 Soln. Date: 07/23/2016 Soln. No.: 00214
 File Date: 07/23/2016

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDWK S3-0386
 Expires: 03/22/2018
 Bottle No.: 0417

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:08D		
Control 1 EC	0.099%	09:09D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10D		
Control 2 EC	0.098%	09:10D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:11D		
Control 3 EC	0.098%	09:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTRALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 07/23/2016