

Calibrating Unit
New Standard Solution Report



ORIGINAL

SOI #216

8/8/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01818
Certification File No.: 01819
Linearity File No.: 01820
Solution File No.: 01917
Sequential File No.: 01917

Serial No.: ARAJ-0075
Calib. Date: 03/18/2016
Cert. Date: 03/18/2016
Lin. Date: 03/18/2016
Soln. Date: 08/08/2016
File Date: 08/08/2016

Serial No.: DDWK S3-0386
Expires: 06/07/2018
Bottle No.: 1209

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16140

Model No.: CU-34

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:52D		
Control 1 EC	0.099%	10:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54D		
Control 2 EC	0.098%	10:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		
Control 3 EC	0.098%	10:55D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 08/08/2016