

Calibrating Unit

New Standard Solution Report

 ORIGINAL

Sol. 218

8-20-16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01818
Certification File No.: 01819
Linearity File No.: 01820
Solution File No.: 01924
Sequential File No.: 01924

Serial No.: ARAJ-0075
Calib. No.: 00021
Cert. No.: 00015
Lin. No.: 00015
Soln. No.: 00218
File Date: 08/20/2016

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 06/07/2018
Bottle No.: 1206

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:48D		
Control 1 EC	0.098%	05:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:50D		
Control 2 EC	0.098%	05:50D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:51D		
Control 3 EC	0.098%	05:51D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:52D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: 

Badge No.: 306

Date: 08/20/2016