

Calibrating Unit New Standard Solution Report

Sol. # 220
10/14/16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01934
Certification File No.: 01935
Linearity File No.: 01936
Solution File No.: 01937
Sequential File No.: 01937

Serial No.: ARAJ-0075

Calib. Date: 10/14/2016
Cert. Date: 10/14/2016
Lin. Date: 10/14/2016
Soln. Date: 10/14/2016
File Date: 10/14/2016

Calib. No.: 00023
Cert. No.: 00016
Lin. No.: 00016
Soln. No.: 00220

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 06/07/2018
Bottle No.: 1313

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:42D		
Control 1 EC	0.100%	13:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:43D		
Control 2 EC	0.099%	13:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:44D		
Control 3 EC	0.099%	13:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔXKP2 - 436 (KA)

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tr. I K W. Q #6704

Badge No.: 6704

Date: 10/14/2016