

Calibrating Unit New Standard Solution Report

501-221

ORIGINAL

10-26-16

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01934
Certification File No.: 01935
Linearity File No.: 01936
Solution File No.: 01945
Sequential File No.: 01945

Serial No.: ARAJ-0075
Calib. No.: 00023
Cert. No.: 00016
Lin. No.: 00016
Soln. No.: 00221
File Date: 10/26/2016

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 03/22/2018
Bottle No.: 0806

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:26D		
Control 1 EC	0.099%	10:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:27D		
Control 2 EC	0.098%	10:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29D		
Control 3 EC	0.099%	10:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: PTJ. R.S. [Signature] FA Job

Badge No.: 306

Date: 10/26/2016