

# Calibrating Unit New Standard Solution Report

501 #  
228  
11/4/16

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01482  
Certification File No.: 01483  
Linearity File No.: 01484  
Solution File No.: 01485  
Sequential File No.: 01485

Serial No.: ARWF-0394

Calib. Date: 11/04/2016    Calib. No.: 00036  
Cert. Date: 11/04/2016    Cert. No.: 00026  
Lin. Date: 11/04/2016    Lin. No.: 00026  
Soln. Date: 11/04/2016    Soln. No.: 00228  
File Date: 11/04/2016

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0040  
Expires: 06/07/2018  
Bottle No.: 0042

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:35S	09:35 Δ KA	
Control 1 EC	0.099%	08:36S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:36S		
Control 2 EC	0.099%	08:37S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:38S		
Control 3 EC	0.098%	08:38S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39S	09:39 Δ KA	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔXKPZ-444 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpr. I K W. Q #6704

Badge No.: 6704

Date: 11/04/2016