

# Calibrating Unit

## New Standard Solution Report

Sol #

222

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01934  
Certification File No.: 01935  
Linearity File No.: 01936  
Solution File No.: 01957  
Sequential File No.: 01957

Serial No.: ARAJ-0075

11/6/16

Calib. Date: 10/14/2016  
Cert. Date: 10/14/2016  
Lin. Date: 10/14/2016  
Soln. Date: 11/06/2016  
File Date: 11/06/2016

Calib. No.: 00023  
Cert. No.: 00016  
Lin. No.: 00016  
Soln. No.: 00222

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 07/06/2018  
Bottle No.: 1381

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:29S		
Control 1 EC	0.099%	15:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30S		
Control 2 EC	0.099%	15:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31S		
Control 3 EC	0.099%	15:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: PH. J. J. #336

Badge No.: 336

Date: 11/06/2016