

# Calibrating Unit New Standard Solution Report

 ORIGINAL

Sol. # 224  
11/28/16

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01934  
Certification File No.: 01935  
Linearity File No.: 01936  
Solution File No.: 01972  
Sequential File No.: 01972

Serial No.: ARAJ-0075  
Calib. Date: 10/14/2016  
Cert. Date: 10/14/2016  
Lin. Date: 10/14/2016  
Soln. Date: 11/28/2016  
File Date: 11/28/2016

Calib. No.: 00023  
Cert. No.: 00016  
Lin. No.: 00016  
Soln. No.: 00224

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 201605E

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 05/17/2018  
Bottle No.: 1215

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:03S		
Control 1 EC	0.098%	09:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04S		
Control 2 EC	0.098%	09:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05S		
Control 3 EC	0.098%	09:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: \_\_\_\_\_

*[Handwritten Signature]* 306

Badge No.: 306

Date: 11/28/2016