

# Calibrating Unit

## New Standard Solution Report

**ORIGINAL**  
*sol. # 229*  
*11/28/16*

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 01482  
 Certification File No.: 01483  
 Linearity File No.: 01484  
 Solution File No.: 01490  
 Sequential File No.: 01490

Serial No.: ARV 0394  
 Calib. No.: 00036  
 Cert. No.: 00026  
 Lin. No.: 00026  
 Soln. No.: 00229  
 File Date: 11/28/2016

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 16140

Model No.: CU-34  
 Serial No.: DDAN-0040  
 Expires: 06/07/2018  
 Bottle No.: 1338

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:44S		
Control 1 EC	0.098%	08:45S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:45S		
Control 2 EC	0.097%	08:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:47S		
Control 3 EC	0.098%	08:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: \_\_\_\_\_

*RWF Job*

Badge No.: 306

Date: 11/28/2016