

Calibrating Unit New Standard Solution Report

SOL# 225
ORIGINAL

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01934
Certification File No.: 01935
Linearity File No.: 01936
Solution File No.: 01979
Sequential File No.: 01979

Serial No.: ARAJ-0075

Calib. Date: 10/14/2016 Calib. No.: 00023
Cert. Date: 10/14/2016 Cert. No.: 00016
Lin. Date: 10/14/2016 Lin. No.: 00016
Soln. Date: 12/05/2016 Soln. No.: 00225
File Date: 12/05/2016

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16140

Model No.: CU-34 Serial No.: DDAN-0041
Expires: 06/07/2018
Bottle No.: 1335

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:10S		
Control 1 EC	0.098%	20:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:11S		
Control 2 EC	0.098%	20:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:12S		
Control 3 EC	0.097%	20:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

FF J06

Badge No.: 306

Date: 12/05/2016