

Calibrating Unit

New Standard Solution Report



ORIGINAL

SOL. 226
12/11/16

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01934
 Certification File No.: 01935
 Linearity File No.: 01936
 Solution File No.: 01984
 Sequential File No.: 01984

Calib. Date: 10/14/2016
 Cert. Date: 10/14/2016
 Lin. Date: 10/14/2016
 Soln. Date: 12/11/2016
 File Date: 12/11/2016

Serial No.: ARAJ-0075
 Calib. No.: 00023
 Cert. No.: 00016
 Lin. No.: 00016
 Soln. No.: 00226

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0041
 Expires: 06/07/2018
 Bottle No.: 1210

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 04:43S | | |
| Control 1 EC | 0.099% | 04:44S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 04:44S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:45S | | |
| Control 2 EC | 0.098% | 04:45S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 04:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:46S | | |
| Control 3 EC | 0.097% | 04:47S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 04:47S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:47S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Robert Westfall

Badge No.: 306

Date: 12/11/2016