

# Calibrating Unit New Standard Solution Report

Sol. # 227  
12-14-16

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 01985 Calib. Date: 12/14/2016 Calib. No.: 00024  
Certification File No.: 01986 Cert. Date: 12/14/2016 Cert. No.: 00017  
Linearity File No.: 01987 Lin. Date: 12/14/2016 Lin. No.: 00017  
Solution File No.: 01988 Soln. Date: 12/14/2016 Soln. No.: 00227  
Sequential File No.: 01988 File Date: 12/14/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041  
Control Solution %: 0.100% Expires: 06/07/2018  
Solution Control Lot: 16140 Bottle No.: 1384

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:32S		
Control 1 EC	0.100%	10:33S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S		
Control 2 EC	0.099%	10:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35S		
Control 3 EC	0.099%	10:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔXK P2-436 (KA)

### Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpr. I K W. Q #6704

Badge No.: 6704

Date: 12/14/2016