

# Calibrating Unit

## New Standard Solution Report

JAN - 3 2017

*Sol # 230*  
*12/18/16*

**Equipment** Alcotest 7110 MKIII-C  
**Location:** TOMS RIVER P.D.  
 Calibration File No.: 01482  
 Certification File No.: 01483  
 Linearity File No.: 01484  
 Solution File No.: 01497  
 Sequential File No.: 01497

Serial No.: ARWF-0394  
 Calib. No.: 00036  
 Cert. No.: 00026  
 Lin. No.: 00026  
 Soln. No.: 00230

Calib. Date: 11/04/2016  
 Cert. Date: 11/04/2016  
 Lin. Date: 11/04/2016  
 Soln. Date: 12/18/2016  
 File Date: 12/18/2016

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0040  
 Expires: 06/07/2018  
 Bottle No.: 1235

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:00S		
Control 1 EC	0.098%	22:01S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:02S		
Control 2 EC	0.098%	22:02S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:03S		
Control 3 EC	0.097%	22:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: \_\_\_\_\_

*PH. In Later #336*

Badge No.: 336

Date: 12/18/2016