



**ORIGINAL**

# Calibrating Unit New Standard Solution Report

Sol # 231  
4/6/17

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01482  
Certification File No.: 01483  
Linearity File No.: 01484  
Solution File No.: 01503  
Sequential File No.: 01503

Calib. Date: 11/04/2016  
Cert. Date: 11/04/2016  
Lin. Date: 11/04/2016  
Soln. Date: 01/06/2017  
File Date: 01/06/2017

Serial No.: ARWF-0394  
Calib. No.: 00036  
Cert. No.: 00026  
Lin. No.: 00026  
Soln. No.: 00231

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16140

Model No.: CU-34

Serial No.: DDAN-0040  
Expires: 06/07/2018  
Bottle No.: 1388

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 14:23S        |                               |                           |
| Control 1 EC      | 0.098%         | 14:24S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.098%         | 14:24S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:25S        |                               |                           |
| Control 2 EC      | 0.097%         | 14:26S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.097%         | 14:26S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:26S        |                               |                           |
| Control 3 EC      | 0.097%         | 14:27S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.097%         | 14:27S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:28S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: Kenneth Thomas #314

Badge No.: 314

Date: 01/06/2017