

Calibrating Unit

New Standard Solution Report **ORIGINAL**

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARAJ-0075 |
| Location: | TOMS RIVER P.D. | |
| Calibration File No.: | 01985 | Calib. Date: 12/14/2016 |
| Certification File No.: | 01986 | Calib. No.: 00024 |
| Linearity File No.: | 01987 | Cert. Date: 12/14/2016 |
| Solution File No.: | 02030 | Cert. No.: 00017 |
| Sequential File No.: | 02030 | Lin. Date: 12/14/2016 |
| | | Lin. No.: 00017 |
| | | Soln. Date: 02/19/2017 |
| | | Soln. No.: 00233 |
| | | File Date: 02/19/2017 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWK S3-0386 |
| Solution Control Lot: | 16140 | Expires: 06/07/2018 |
| | | Bottle No.: 1211 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 14:15S | | |
| Control 1 EC | 0.098% | 14:16S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 14:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:16S | | |
| Control 2 EC | 0.097% | 14:17S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 14:17S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:18S | | |
| Control 3 EC | 0.097% | 14:18S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 14:18S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:19S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 02/19/2017