

# Calibrating Unit New Standard Solution Report

SOL # 234  
2/26/17

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01985  
Certification File No.: 01986  
Linearity File No.: 01987  
Solution File No.: 02034  
Sequential File No.: 02034

Serial No.: ARAJ-0075

Calib. Date: 12/14/2016  
Cert. Date: 12/14/2016  
Lin. Date: 12/14/2016  
Soln. Date: 02/26/2017  
File Date: 02/26/2017

Calib. No.: 00024  
Cert. No.: 00017  
Lin. No.: 00017  
Soln. No.: 00234

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16390

Model No.: CU-34

Serial No.: DDWK S3-0386  
Expires: 12/02/2018  
Bottle No.: 1272

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:29S		
Control 1 EC	0.098%	00:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30S		
Control 2 EC	0.098%	00:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:31S		
Control 3 EC	0.099%	00:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:

 # 350

Badge No.: 350

Date: 02/26/2017