



ORIGINAL

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.	Calib. Date:	11/04/2016
Calibration File No.:	01482	Calib. No.:	00036
Certification File No.:	01483	Cert. Date:	11/04/2016
Linearity File No.:	01484	Lin. Date:	11/04/2016
Solution File No.:	01567	Soln. Date:	03/31/2017
Sequential File No.:	01567	File Date:	03/31/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1351

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:51S		
Control 1 EC	0.099%	23:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:53S		
Control 2 EC	0.098%	23:53S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:54S		
Control 3 EC	0.099%	23:55S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: THOMAS

First Name: KENNETH

MI: J

Signature: _____

Kenneth Thom #314

Badge No.: 314

Date: 03/31/2017