

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02042 Calib. Date: 04/18/2017 Calib. No.: 00026
Certification File No.: 02043 Cert. Date: 04/18/2017 Cert. No.: 00018
Linearity File No.: 02044 Lin. Date: 04/18/2017 Lin. No.: 00018
Solution File No.: 02045 Soln. Date: 04/18/2017 Soln. No.: 00237
Sequential File No.: 02045 File Date: 04/18/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 12/02/2018
Solution Control Lot: 16390 Bottle No.: 1207

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:40D		
Control 1 EC	0.101%	11:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		
Control 2 EC	0.100%	11:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43D		
Control 3 EC	0.100%	11:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJPZ-025 **(KA)**

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TPF-I KW A #6704

Badge No.: 6704

Date: 04/18/2017