Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARWF-0394
Location: TOMS RIVER P.D.

Calibration File No.:	01582	Calib. Date:	04/18/2017	Calib. No.:	00037
Certification File No.:	01583	Cert. Date:	04/18/2017	Cert. No.:	00027
Linearity File No.:	01584	Lin. Date:	04/18/2017	Lin. No.:	00027
Solution File No.:	01585	Soln. Date:	04/18/2017	Soln. No.:	00240
Sequential File No.:	01585	File Date:	04/18/2017		

Calibrating Unit: WET — Model No.: CU-34 — Serial No.: DDWK S3-0387
Control Solution %: 0.100% — Expires: 12/02/2018
Solution Control Lot: 16390 — Bottle No.: 1350

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:41D		
Control 1 EC	0.100%	13:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42D		
Control 2 EC	0.100%	13:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:44D		
Control 3 EC	0.099%	13:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: ΔΔWJPZ-026 (KA)

Changed By:
Last Name: ALCOTT — First Name: KEVIN — MI: W.

Signature: Tpr. I K W. [illegible] #6704 — Badge No.: 6704
Date: 04/18/2017