

Item 14

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02054	Calib. Date:	05/12/2017
Certification File No.:	02055	Cert. No.:	00028
Linearity File No.:	02056	Cert. Date:	05/12/2017
Solution File No.:	02057	Lin. No.:	00019
Sequential File No.:	02057	Soln. Date:	05/12/2017
		Soln. No.:	00240
		File Date:	05/12/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1221

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:18D		
Control 1 EC	0.101%	14:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:19D		
Control 2 EC	0.101%	14:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:20D		
Control 3 EC	0.101%	14:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DAWJ PZ-025 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpr. I K W. #6704

Badge No.: 6704

Date: 05/12/2017